



Simec AG

Areal Bleiche West Postfach 413 CH - 4800 Zofingen
Telefon +41 62 752 83 08 Fax +41 62 752 83 09
E-Mail: info@simec.ch www.simec.ch

Test report

Number **C 42786**

Object of analysis: **Biomex Equipoise 200**

Customer:

Samples: **1**

Batch: **Q2268V954, Exp. 06/2017**

Receipt: **07.04.2016**

Subject: **Boldenone undecylenate: Assay, Identification**

Method(s): **HPLC**

Sample / Analysis:	Method	Status	Result	Specification	complies
<u>Boldenone undecylenate:</u>					
Identification	HPLC	N	complies	complies	yes
Content	HPLC	N	223.9 mg/mL	200 mg/mL	-

Method-status: G = GMP A = accredited V = generally validated P = validated on product N = not validated
E = external lab

Remarks: The sample(s) mentioned above have been analysed as they have been sent to us by the client. We have not controlled external processes like manufacturing, labelling, sampling, shipping, storage. These results are for information only and do not compensate for correct quality control by the manufacturer/distributor. Generally, pharmaceutical products have to be produced and distributed under full GMP/GDP regime, including GMP-compliant analyses with validated methods.

Signatures:
controlled & released:

Staff:


completed:
28.04.2016

The enclosed results refer exclusively to the object of examination described above. The measuring accuracy is available on request. Only handwritten signatures are valid. Without written permission, it is not allowed to publish single parts of this report.

page 1 of 1

Report steroid analysis

HPLC:	Nr.5	Druckdatum:	28.04.2016 08:23:30
Analysendatum:	27.04.2016 10:52:33	Injektvolumen:	20
Vial:	61		
Probenidentifikation:	STD Boldenone Undecylenat		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



Channel A

UV Results


Name	Time	Area	Amount	Area Percent
	21,657	3310594	0,000	2,345
	21,913	2881625	0,000	2,041
Boldenone undecylenate	35,153	134974620	0,000	95,614

Totals		141166839	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep1.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 27.04.2016 11:46:06 Druckdatum: 28.04.2016 08:23:35
 Vial: 61 Injektvolumen: 20
 Probenidentifikation: STD Boldenone Undecylenat
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm




Channel A

UV Results				
Name	Time	Area	Amount	Area Percent
	21,927	3350316	0,000	2,406
	22,177	2781031	0,000	1,997
Boldenone undecylenate	35,277	133117405	0,000	95,597
Totals		139248752	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep2.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC:	Nr.5	Druckdatum:	28.04.2016 08:23:39
Analysendatum:	27.04.2016 12:39:37	Injektvolumen:	20
Vial:	61		
Probenidentifikation:	STD Boldenone Undecylenat		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



Channel A

UV Results


Name	Time	Area	Amount	Area Percent
	21,903	3362705	0,000	2,405
	22,153	2782216	0,000	1,990
Boldenone undecylenate	35,230	133679338	0,000	95,605

Totals		139824259	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep3.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC:	Nr.5	Druckdatum:	28.04.2016 08:23:43
Analysendatum:	27.04.2016 13:33:08	Injektvolumen:	20
Vial:	61		
Probenidentifikation:	STD Boldenone Undecylenat		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



Channel A


UV Results

Name	Time	Area	Amount	Area Percent
	21,907	3352131	0,000	2,406
	22,157	2762350	0,000	1,983
Boldenone undecylenate	35,253	133180622	0,000	95,610
Totals		139295103	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep4.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC:	Nr.5	Druckdatum:	28.04.2016 08:23:47
Analysendatum:	27.04.2016 14:26:37	Injektvolumen:	20
Vial:	61		
Probenidentifikation:	STD Boldenone Undecylenat		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



Channel A

UV Results


Name	Time	Area	Amount	Area Percent
	21,900	3343139	0,000	2,401
	22,150	2762428	0,000	1,984
Boldenone undecylenate	35,210	133123646	0,000	95,615

Totals		139229213	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep5.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC:	Nr.5	Druckdatum:	28.04.2016 08:23:52
Analysendatum:	27.04.2016 15:20:07	Injektvolumen:	20
Vial:	61		
Probenidentifikation:	STD Boldenone Undecylenat		
Operator:	admin		
Chrom.-Beding.	ACN / Wasser, gradient		
Temperatur:	40 °C		
Säule	Nr. 121, 250 x 4,6, C18		
Detektion:	UV 240nm		



Channel A


UV Results				
Name	Time	Area	Amount	Area Percent
	21,913	3355160	0,000	2,409
	22,163	2744010	0,000	1,970
Boldenone undecylenate	35,253	133189094	0,000	95,621
Totals		139288264	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\082_STD Boldenone Undecylenat-Rep6.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 27.04.2016 22:27:34
 Vial: 64
 Probenidentifikation: C42786 A
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm

Druckdatum: 28.04.2016 08:24:25
 Injektvolumen: 20



Channel A

UV Results

Name	Time	Area	Amount	Area Percent
	16,353	56731799	0,000	35,662
Boldenone undecylenate	35,213	101295600	0,000	63,675
	40,513	1055136	0,000	0,663


Totals		159082535	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\086_C42786 A-Rep1.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 27.04.2016 23:20:59
 Vial: 64
 Probenidentifikation: C42786 A
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm

Druckdatum: 28.04.2016 08:24:29
 Injektvolumen: 20



Channel A

UV Results

Name	Time	Area	Amount	Area Percent
	16,340	56728254	0,000	35,704
Boldenone undecylenate	35,270	101101281	0,000	63,632
	40,587	1055407	0,000	0,664


Totals		158884942	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\086_C42786 A-Rep2.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 28.04.2016 00:14:26
 Vial: 64
 Probenidentifikation: C42786 A
 Operator: admin
 Chrom.-Beding.: ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule: Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm

Druckdatum: 28.04.2016 08:24:34
 Injektvolumen: 20



Channel A

UV Results


Name	Time	Area	Amount	Area Percent
	16,377	56843715	0,000	35,692
Boldenone undecylenate	35,277	101363292	0,000	63,646
	40,587	1053774	0,000	0,662

Totals		159260781	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\086_C42786 A-Rep3.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 28.04.2016 01:07:52 Druckdatum: 28.04.2016 08:24:38
 Vial: 65 Injektvolumen: 20
 Probenidentifikation: C42786 B
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm



Channel A

UV Results

Name	Time	Area	Amount	Area Percent
	16,370	59528315	0,000	35,506
Boldenone undecylenate	35,267	107017711	0,000	63,832
	40,580	1110006	0,000	0,662


Totals		167656032	0,000	100,000
--------	--	-----------	-------	---------

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse
 Steroide\19,04,16\087_C42786 B-Rep1.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 28.04.2016 02:01:17
 Vial: 65
 Probenidentifikation: C42786 B
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm

Druckdatum: 28.04.2016 08:24:41
 Injektvolumen: 20




Channel A

UV Results				
Name	Time	Area	Amount	Area Percent
	16,387	59481479	0,000	35,574
Boldenone undecylenate	35,247	106611701	0,000	63,760
	40,557	1113333	0,000	0,666
Totals		167206513	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse Steroide\19,04,16\087_C42786 B-Rep2.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met

Report steroid analysis

HPLC: Nr.5
 Analysendatum: 28.04.2016 02:54:33 Druckdatum: 28.04.2016 08:24:45
 Vial: 65 Injektvolumen: 20
 Probenidentifikation: C42786 B
 Operator: admin
 Chrom.-Beding. ACN / Wasser, gradient
 Temperatur: 40 °C
 Säule Nr. 121, 250 x 4,6, C18
 Detektion: UV 240nm



Channel A

UV Results

Name	Time	Area	Amount	Area Percent
	16,377	59633816	0,000	35,622
Boldenone undecylenate	35,217	106668653	0,000	63,718
	40,513	1104015	0,000	0,659
Totals		167406484	0,000	100,000

Daten File: C:\EZChrom Elite\Enterprise\Projects\Simec\Data\2016\April 2016\Diverse
 Steroide\19,04,16\087_C42786 B-Rep3.dat
 Methode: C:\EZChrom Elite\Enterprise\Projects\Simec\Method\Steroide\Steroide A.met